

---

### Lead Equivalencies

---

	100 kVp	120 kVp	130 kVp
Lead Equivalency (mm Pb eq.)			
0.25 Apron	0.34	0.32	0.30
0.35 Apron	0.44	0.41	0.40
0.50 Apron	0.53	0.50	0.50
Transmission (%)			
0.25 Apron	7.40	10.10	12.10
0.35 Apron	4.93	6.80	8.14
0.50 Apron	3.45	4.62	5.53

---

Tested by Massachusetts General Hospital